

0000 7808



WASHINGTON STATE PATROL SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: 4/20/2006 INITIALS: Dec OWNER: WSP
 DISTRICT: 2 ADDRESS: _____
 LOCATION: 7 M Shop COUNTY: _____
 MAKE: Hoenni TYPE: port wheel weighing ACCURACY CLASS: III
 MODEL: 10' ID. NO: 27597 CAPACITY: 20,000 lbs SECTIONS: 1
 MINIMUM GRADUATIONS: 50 REG. ELEMENT: clial PLATFORM: 15" X 26"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
Center	5000	5000	0				
"	10000	10000	0				
"	15000	15000	0				
Center	20000	20000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 19577

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cronin Date: 4/26/2006

State of Washington
 County of Thurston
 Signed and sworn to (or affirmed) before
 me on 4-20-06 by

Michael J. Evans
 (Signature of notary public)
Michael J. EVANS

Printed Name of Notary 12-27-07
 My appointment expires



3000-362-021 (R) (10/98)

DISTRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS